

Convolutional Neural Networks and AI Technology for Early Detection of Lung Cancer: A Deep Learning Approach

Ganashree TS^{#*}, Mesith Chaimanee, Mithileysh Sathiyarayanan

Shinawatra University, Faculty of Engineering, Thailand. [#]Post-Doctoral Research Scholar. ^{*}Corresponding Author's Email: ganashree2024@gmail.com

Abstract

Lung cancer detection using Convolutional Neural Networks (CNNs) represents a transformative approach in medical imaging, utilizing deep learning to identify cancerous images in lung images. Our project aims to develop and implement a CNN-based system to detect lung cancer from computed tomography (CT) scans. The process begins with preprocessing the CT images to standardize and highlight key features, followed by training a multi-layer CNN designed to recognize complex visual images. For this project, the key technologies used include Python for scripting, Tensor Flow as the deep learning framework, and CPU acceleration to expedite model training. The CNN architecture consists of layers for convolution, activation functions, pooling, and dropout to reduce over fitting. The model is trained on a sizable dataset of labeled lung images, comprising a diverse mix of normal and cancerous cases, providing a solid foundation for robust learning. The final results indicate a high accuracy of 95%, signifying a highly effective model with low rates of false positives and false negatives. This level of accuracy demonstrates that the CNN-based approach can substantially support radiologists in early lung cancer detection, which could lead to improved patient outcomes through earlier diagnosis and intervention. The success of this system underscores the potential of deep learning to transform medical diagnostics and assist healthcare professionals in making more informed clinical decisions.

Keywords: Computed Tomography, Convolutional Neural Networks, Deep Learning, Lung Cancer Detection, Model Accuracy.

Introduction

Lung cancer detection through CNNs represents a cutting-edge approach to medical imaging. Due to the high mortality rates associated with lung cancer, early detection is crucial for enhancing patient survival rates. This paper explores the development of a CNN-based system to detect lung cancer from CT scans, emphasizing accuracy and robustness (1). Lung cancer detection using CNNs is a new method in medical imaging, driven by the need for early and accurate diagnosis of this deadly disease. Lung cancer could be considered one of the chief causes of deaths from cancer globally, though; early detection can significantly improve survival rates (2). A CNN-focused system for detecting lung cancer in CT scans, with a goal to improve diagnostic accuracy and ease the workload on healthcare professionals, is explored in this paper. CNNs are ideal for this project because of their automatic features that learn and extract complex features from images (3). In our approach, we preprocess CT scans to standardize

the data and emphasize key features, then train a multi-layer CNN model to find patterns that indicate lung cancer. This deep learning model is trained on a large dataset of labeled lung images, including both cancerous and non-cancerous cases, providing a solid base for learning (4). Our aim is to tackle key challenges in lung cancer detection, such as reducing false positives and negatives, and ensuring the system works across different groups of people and various stages of cancer (5). The high accuracy rates achieved by our CNN model show that deep learning has the potential to transform lung cancer diagnostics. The research results indicate this method is able to help radiologists come up with diagnoses with higher accuracy, following up earlier treatment and better patient outcomes (6). Overall, this research is a meaningful step in the expanding field of medical AI, offering a promising solution to one of healthcare's toughest problems (7).

This is an Open Access article distributed under the terms of the Creative Commons Attribution CC BY license (<http://creativecommons.org/licenses/by/4.0/>), which permits unrestricted reuse, distribution, and reproduction in any medium, provided the original work is properly cited.

(Received 10th January 2025; Accepted 10th April 2025; Published 28th April 2025)

Key Contributions in this article are listed below:

- Introduction of a robust CNN model tailored for lung cancer detection.
- Demonstrate high accuracy, validating deep learning as a valuable tool for medical diagnostics (8).
- Use of a comprehensive dataset with both normal and cancerous cases, allowing for extensive testing and validation.

This article is systematized in the following manner. The "Methodology" segment details on CNN architecture and the pre-processing techniques used to prepare the CT images. The "Experimentation" section describes the dataset, training processes, hyper parameters, and computational resources employed. The "Results and Discussion" section examines the model's accuracy, evaluates its performance, and discusses its impact on early lung cancer detection. Finally, the "Conclusion" section summarizes the key findings and outlines future research directions to enhance lung cancer detection using CNNs. In the field of lung cancer detection, several traditional methods have been utilized to analyze lung images, with Support Vector Machines (SVMs) and Local Binary Patterns (LBPs) among the most prominent. SVMs have been widely used for their ability to classify complex datasets by finding cancers. LBPs, a texture-based feature extraction technique, have been used to quantify the patterns in lung tissue, providing valuable data for classification algorithms like SVM. Moreover, these techniques might struggle with varying image quality or diverse patient demographics, leading to reduced accuracy and reliability. Our method proposes using "CNNs" for acknowledging regarding "SVMs" and "LBPs" limitations. CNNs are proficient in automatic extraction and learning capacities regarding complex aspects of unprocessed pictures of lungs excluding the necessity for manual interferences (9). This method of deep learning holds the ability to augment the detection of lung cancer with precision and decrease false positives and false negatives incidences. For further enhancements over the pre-existing techniques, this projected system comprises an exclusive "CNN architecture" planned precisely for the purpose of lung cancer detection, joined by an effectual "pre-processing pipeline" regulating CT scan images. This combination allows the model to detect cancerous

with a high degree of accuracy and robustness. Looking ahead, future research could explore hybrid models that combine the strengths of SVMs with the deep learning capabilities of CNNs, potentially creating more versatile systems. Besides, adding extra data sources, counting "patient history" and "genetic information", to accelerate the meticulousness and dependability of the detection process. This study might in due course guide on the way to a further operative and capable recognition device for lung cancer (10-12). Current lung cancer detection technologies confront numerous distinct problems, which might impair the accuracy, efficiency, and timeliness of diagnosis. These difficulties include: Subjectivity and Inter-observer Variability: Radiological interpretation of CT images strongly relies on the competence of the radiologist. This may lead to high inter-observer variability, as different radiologists may interpret the same picture differently, possibly leading to misinterpretation or missing lesions. Late Diagnosis and Poor Prognosis: Lung cancer is typically identified at a late stage when the illness has already advanced, leading to a worse prognosis. This is largely owing to the subtle and frequently asymptomatic character of early-stage lung cancer, which may be difficult to identify by conventional approaches like chest X-rays or manual CT scan interpretation. Difficulty in Detecting Small or Subtle Lesions: Small lesions, especially those in early-stage lung cancer, may be tough to detect with traditional approaches. Tumors that appear as tiny nodules or that are situated in difficult-to-reach parts of the lung may be neglected owing to the resolution limits of imaging or the difficulty of differentiating benign from malignant growths. High False Positive and False Negative Rates: Traditional procedures, such as chest X-rays and even early CT scans, may result in high rates of false positives (non-cancerous lesions wrongly labeled as cancer) and false negatives (malignant lesions missed totally). These incorrect findings might lead to needless biopsies and therapies or delayed diagnosis in the event of false negatives. Limited Capacity for Large Data Handling: As the number and resolution of imaging data expand, conventional techniques for processing, including manual review by radiologists, become less efficient. Large datasets, particularly from high-resolution CT scans, are difficult to evaluate

properly in a short period of time, which slows down the diagnostic and decision-making process.

Methodology

Generating a “CNN” for sensing lung cancer necessitates a “systematic approach”. This methodology outlines the critical steps, from gathering data to implementing the model, ensuring a reliable and accurate system for detecting lung cancer in CT scans. The proposed methodology for lung cancer detection using CNNs with lung images starts with data collection, sourcing a diverse set of CT scans from publicly available datasets or medical institutions. Next, the data is pre-processed by resizing images to a uniform size, and converting them to gray scale or

normalizing them to ensure consistency (13). After pre-processing, the data is visualized to detect any anomalies or inconsistencies, ensuring the dataset is clean and reliable (14-18). The final step involves implementing the CNN model, typically using layers for convolution, activation (like ReLU), pooling, and dropout to prevent over fitting. “Training set” is used for model training, fine-tuned by means of the “validation set”, and then assessed through the “test set” to measure its correctness. The following systematic approach has been designed to create a robust and accurate structure for the purpose of lung cancer detection (19). The below Figure 1 shows the Architecture of proposed methodology.




Figure 1: Architecture of Proposed Methodology




Figure 2: Architecture of Machine Learning Model




Figure 3: Architecture of CNN




Figure 4: ANN Architecture

Data Collection

The first step is to collect CT scan images of the lungs, with a mixture of healthy and cancerous cases. To ensure a diverse dataset, consider the distribution across various patient demographics, cancer stages, and image quality (20). The Figure 2 shows the ANN Architecture, the Figure 3 shows the Architecture of CNN and Figure 4 details the ANN architecture.

Data Pre-Processing

Resizing and Normalization

CT images vary in size, so they are resized to a standard format, typically 224×224 pixels. The resizing operation involves interpolation, often done with techniques like bicubic or bilinear interpolation. Mathematically, resizing is a mapping operation from an original set of pixel coordinates to a new set which is given in eqn. [1] and [2]:

$$I'(x, y) = I(x \times WW_0, y \times HH_0) \quad [1]$$

$$I'(x, y) = I(W_0x \times W, H_0y \times H) \quad [2]$$

Where I' is the resized image, I is the original image, W_0 and H_0 are the original width and height, and W and H are the target width and height. "Normalization scales" the "pixel values" to a range of $[0, 1]$, helping in augmenting convergence during training. This is calculated as eqn. [3] and [4]:

$$I_{norm}(x, y) = \frac{I(x, y) - \min(I)}{\max(I) - \min(I)} \quad [3]$$

$$I_{norm}(x, y) = \frac{I(x, y) - \min(I)}{\max(I) - \min(I)} \quad [4]$$

where $I(x, y)$ are the original pixel values, and $\max(I)$ and $\min(I)$ denote the maximum and minimum pixel values in the image.

Data Visualization

Using tools like Matplotlib or Seaborn, a sample of the preprocessed data is visualized to check for anomalies or inconsistencies. This involves generating plots or displaying images to visually inspect the data quality.

Data Splitting and Classification

Data splitting notes the procedure of separating and distributing the dataset among "training",

“validation”, and “test” sets. The typical ratio is 80/10/10. To create these subsets, random sampling is often used to ensure each subset is representative of the whole dataset. The “training set” is employed for training the CNN model, the “validation set” is integrated for tuning hyperparameters, and the “test set” is used for final model evaluation (21).

Implementation of the CNN Model

The CNN model consists of various layers. Convolutional layers apply filters to extract features from the images. The “convolution operation” can be represented mathematically as eqn. [5]:

$$F(x, y) = \sum_{i=0}^{k-1} \sum_{j=0}^{k-1} W(i, j) \cdot I(x + i, y + j) \quad [5]$$

Where $F(x, y)$ denotes the output feature map, $W(i, j)$ indicates convolutional filter, $I(x, y)$ is used for the input image, and k is the kernel size (22). Pooling layers are implemented for spatial dimensions of the feature maps reduction, often employing “max pooling” in eqn. [6] (23),

$$P(x, y) = i, j \max \{F(x + i, y + j)\} \quad [6]$$

where $P(x, y)$ is the pooled output.

Randomly setting a proportion of activations to zero through “Dropout layers” during training to avoid overfitting, with the dropout probability p typically ranging from 0.2 to 0.5.

Model Training and Evaluation

The “CNN model” is trained through the training set, with hyperparameters tuned on the basis of

the performance of the validation set (24). Usual “hyperparameters” involve the “number of epochs”, “learning rate”, and “batch size”. During training, the “binary cross-entropy loss function” is used in eqn. [7].

$$L(y, \hat{y}) = -y \cdot \log(\hat{y}) - (1 - y) \cdot \log(1 - \hat{y}) \quad [7]$$

where y represents the true label, and $\hat{y}$ represents the forecasted probability for lung cancer. The final model is evaluated using the test set to measure accuracy, precision, and recall. Accuracy is calculated in the following manner in eqn. [8], [7], (8, 20)

$$Accuracy = \frac{TP+TN}{TP+TN+FP+FN} \quad [8]$$

With these mathematical operations and calculations, the proposed methodology for lung cancer detection using CNNs can be implemented to effectively classify lung cancer in CT scans (25). Finally, give any input image and then do pre-process for it after that compare input features with database features then u will get those pictures having lung cancer or lung health. This methodology provides a clear and straightforward approach to building a CNN-based lung cancer detection system, emphasizing data quality, model design, and proper evaluation to ensure the final product is robust and accurate. The below Table 1 shows the epochs with respective loss and accuracy levels (randomly 10 values) for CNN model.

Table 1: Tabular Representation of Epochs with Respective Loss and Accuracy Levels (randomly 10 values) for the ANN model

Epoch	Loss	Accuracy	Val_Loss	Val_Accuracy
1/50	1.9625e-05	1.0000	1.8564e-05	1.0000
2/50	1.7703e-05	1.0000	1.7404e-05	1.0000
3/50	1.7595e-05	1.0000	1.6371e-05	1.0000
4/50	1.6066e-05	1.0000	1.5329e-05	1.0000
5/50	1.5332e-05	1.0000	1.4433e-05	1.0000
46/50	3.0826e-06	1.0000	2.8956e-06	1.0000
47/50	2.7987e-06	1.0000	2.7534e-06	1.0000
48/50	2.8094e-06	1.0000	2.6705e-06	1.0000
49/50	2.6410e-06	1.0000	2.5596e-06	1.0000
50/50	2.5944e-06	1.0000	2.4822e-06	1

Results and Discussion

In the first step, we have to assign the path for the dataset using “operating system for cancer” and without tumor cases. “Theos. Path” module is an authentic significantly implemented module handy during recycling lines from multiple places within

the system. It's employed for different purposes similar as for incorporating, homogenizing along with reacquiring “path names” in “python”. As parameters, these functions acknowledge only bytes or only string objects singularly. It is being

run on Zilches and the results are specific to that. The Figure 5 shows the graphs accuracy details.

CNN Model Output Screenshot

As per Figure 5 graph discussion between accuracy will be done for CNN with data set training more than 95 percent accuracy. In the Figure 6 graph discussion between losses will be done for CNN

with data set training more than less percent accuracy. After training the model, when selected input image from test images then the below Figure 6 shows the result of abnormal means person having the lung cancer done by CNN model with more accurate result. The Figure 7 shows the abnormal lung cancer.




Figure 5: Graphs for Accuracy Details for CNN




Figure 6: Graphs for Loss for CNN




Figure 7: Lung Cancer Abnormal for CNN




Figure 8: Applied OTSU Thresholding1 for Lung Cancer Image for CNN




Figure 9: Applied OTSU Thresholding2 for Lung Cancer Image for CNN

In results, Figure 8 saying after completion of detection lung cancer then it should identify the area using OTSU thresholding 1 as per above Figure 8. If result normal it doesn't show any thresholding's. In results, Figure 9 saying after completion of detection lung cancer then it should identify the area using OTSU thresholding2 as per above Figure 9. If result normal it doesn't show any thresholding's. The below Table 2 shows the F1 score, precision, Specificity details.

ANN Output Screenshot

As per Figure 10 graph discussion between accuracy will be done for ANN with data set training more than 94 percent accuracy. In this Figure 11 graph discussion between losses will be done for ANN with data set training more than less percent accuracy.

Table 2: The Metrics

S. No	Metrics	Actual Results
1	F1 score	0.95
2	Precision	0.94
3	Specificity	0.97




Figure 10: Graphs for Accuracy for ANN




Figure 11: Graphs for Loss for ANN




Figure 12: Lung Cancer Abnormal for ANN




Figure 13: Applied OTSU Thresholding1 for Lung Cancer Image for ANN




Figure 14: Applied OTSU Thresholding2 for Lung Cancer Image for ANN

After training the model, when selected input image from test images then the above Figure 12 shows the result of abnormal means person having the lung cancer done by ANN model with more accurate result. In results, Figure 13 saying after completion of detection lung cancer then it should identify the area using OTSU thresholding1 as per above Figure 13. If result normal it doesn't show any thresholding's. In results, Figure 14 saying

after completion of detection lung cancer then it should identify the area using OTSU thresholding2 as per above Figure 14. If result normal it doesn't show any thresholding's.

Random Forest Classifier Algorithm Output Screenshot

In Figure 15 we applied random forest algorithm for lung dataset images then we got accuracy is 98.36%.

```
# Initialize and train the Random Forest classifier
rf_classifier = RandomForestClassifier(n_estimators=100, random_state=42)
rf_classifier.fit(X_train, y_train)

# Predict on the test set
predictions = rf_classifier.predict(X_test)

# Calculate accuracy
accuracy = accuracy_score(y_test, predictions)
print(f"Accuracy: {accuracy * 100:.2f}%")

Accuracy: 98.36%
```

Figure 15: Training Dataset with Random Forest Algorithm

```
# Initialize the AdaBoost classifier
adaboost_clf = AdaBoostClassifier(n_estimators=50, random_state=42)

# Train the AdaBoost classifier
adaboost_clf.fit(X_train, y_train)

# Make predictions on the test set
y_pred = adaboost_clf.predict(X_test)

# Calculate accuracy
accuracy = accuracy_score(y_test, y_pred)
print(f"Accuracy: {accuracy * 100:.2f}%")

Accuracy: 85.50%
```

Figure 16: Training Dataset with Adaboost Classifier Algorithm




Figure 17: Comparison of Random Forest and Adaboost Classifier

Adaboost Classifier Algorithm Output Screenshot

In Figure 16 we applied Adaboost classifier algorithm for lung dataset images then we got accuracy is 85.50% (9).

Graphical Comparison of Lung Cancer Detection Machine Learning Algorithms

Finally, in Figure 17 shows the comparison of random forest with 98.36% and adaboost classifier with 85.50% for lung cancer detection.

Conclusion

In conclusion, our work on detecting lung cancer through the implementation of “CNNs” with lung images has demonstrated that deep learning can be an effective tool for early diagnosis. The CNN-based approach, combined with a thorough data preprocessing pipeline, achieved high accuracy in detecting cancerous patterns in CT scans. This accuracy level indicates the potential of such systems to assist healthcare professionals and radiologists in making more informed conclusions, eventually steering to earlier interventions and enhanced patient outcomes. There are several areas for future work that could further enhance lung cancer detection. One direction involves exploring hybrid models that combine the strengths of deep learning architectures with generation process, allowing for more flexibility and potentially improved accuracy. Another area of interest is transfer learning, which can leverage pre-trained models for reducing training time and enhance performance through smaller datasets. Integrating supplementary data sources, including clinical records, “genetic information”, even radiomics, could facilitate a highly inclusive understanding into each case, further enhancing the model's exactness. Additionally, explainable AI techniques could be implemented to make the model's predictions more transparent, helping radiologists understand the basis of the algorithm's decisions. These future advancements could lead to more robust and trustworthy systems for lung cancer detection, ultimately revolutionizing medical diagnostics in oncology.

Abbreviation

None.

Acknowledgement

None.

Author Contributions

All authors contributed equally.

Conflict of Interest

The authors have expressed no conflict of interest.

Ethics Approval

Not applicable.

Funding

The authors did not receive support from any organization for the submitted work.

References

1. Ren Z, Zhang Y, Wang S. LCDAE: data augmented ensemble framework for lung cancer classification. *Technology in Cancer Research & Treatment*. 2022 Sep;21(5):15330338221124372.
2. Bhatt D, Patel C, Talsania H, Patel J, Vaghela R, Pandya S, Modi K, Ghayvat H. CNN variants for computer vision: History, architecture, application, challenges and future scope. *Electronics*. 2021 Oct 11;10(20):2470.
3. González-Alday R, García-Cuesta E, Kulikowski CA, Maojo V. A scoping review on the progress, applicability, and future of explainable artificial intelligence in medicine. *Applied Sciences*. 2023 Sep 28;13(19):10778.
4. Davri A, Birbas E, Kanavos T, Ntritsos G, Giannakeas N, Tzallas AT, Batistatou A. Deep learning for lung cancer diagnosis, prognosis and prediction using histological and cytological images: a systematic review. *Cancers*. 2023 Aug 5;15(15):3981.
5. Clark N, Cooper S. Automated lung cancer detection in CT scans using deep learning. *International Journal of Medical Informatics*. 2023; 156:104-114.
6. Patrício C, Neves JC, Teixeira LF. Explainable deep learning methods in medical image classification: A survey. *ACM Computing Surveys*. 2023 Oct 21;56(4):1-41.
7. Rajpal S, Rajpal A, Agarwal M, Kumar V, Abraham A, Khanna D, Kumar N. XAI-CNVMarker: Explainable AI-based copy number variant biomarker discovery for breast cancer subtypes. *Biomedical Signal Processing and Control*. 2023 Jul 1;84:104979.
8. Rahman S, Rahman MM, Faezipour M, Rastgaar M, Ridelman E, Klein JD, Angst BA, Shanti CM. Enhancing Burn Severity Assessment With Deep Learning: A Comparative Analysis and Computational Efficiency Evaluation. *IEEE Access*. 2024 Oct 8;12:147249 - 147268.
9. Hao L, Huang G. An improved AdaBoost algorithm for identification of lung cancer based on electronic nose. *Heliyon*. 2023 Mar 1;9(10):e13633.
10. Gu H, Gan W, Zhang C, Feng A, Wang H, Huang Y, Chen H, Shao Y, Duan Y, Xu Z. A 2D-3D hybrid convolutional neural network for lung lobe auto-segmentation on standard slice thickness computed tomography of patients receiving radiotherapy. *BioMedical Engineering OnLine*. 2021 Dec;20:1-3.

11. Huang D, Li Z, Jiang T, Yang C, Li N. Artificial intelligence in lung cancer: current applications, future perspectives, and challenges. *Frontiers in Oncology*. 2024 Dec 23;14:1486310.
12. Azizi A, Azizi M. Artificial Intelligence Techniques in Medical Imaging: A Systematic Review. *International Journal of Online & Biomedical Engineering*. 2023 Dec 6;19(17):66-97.
13. Thanoon MA, Zulkifley MA, Mohd Zainuri MA, Abdani SR. A review of deep learning techniques for lung cancer screening and diagnosis based on CT images. *Diagnostics*. 2023 Aug 8;13(16):2617.
14. Manju BR, Athira V, Rajendran A. Efficient multi-level lung cancer prediction model using support vector machine classifier. In *IOP Conference Series: Materials Science and Engineering* 2021;1012(1):012034. IOP Publishing.
15. Han R, Yoon H, Kim G, Lee H, Lee Y. Revolutionizing medicinal chemistry: the application of artificial intelligence (AI) in early drug discovery. *Pharmaceuticals*. 2023 Sep 6;16(9):1259.
16. Siddiqi R, Javaid S. Deep learning for pneumonia detection in chest x-ray images: A comprehensive survey. *Journal of imaging*. 2024 Jul 23;10(8):176.
17. Mubarak AS, Serte S, Al-Turjman F, Ameen ZS, Ozsoz M. Local binary pattern and deep learning feature extraction fusion for COVID-19 detection on computed tomography images. *Expert Systems*. 2022 Mar;39(3):e12842.
18. Zheng Y, Wu K, Li J, Tang K, Shi J, Wu H, Jiang Z, Wang W. Partial-label contrastive representation learning for fine-grained biomarkers prediction from histopathology whole slide images. *IEEE Journal of Biomedical and Health Informatics*. 2023;27(12):6144-6266.
19. Prisciandaro E, Sedda G, Cara A, Diotti C, Spaggiari L, Bertolaccini L. Artificial neural networks in lung cancer research: a narrative review. *Journal of Clinical Medicine*. 2023 Jan 22;12(3):880.
20. Patrício C, Neves JC, Teixeira LF. Explainable deep learning methods in medical image classification: A survey. *ACM Computing Surveys*. 2023 Oct 21;56(4):1-41. 21.
21. Wolf AM, Oeffinger KC, Shih TY, Walter LC, Church TR, Fontham ET, Elkin EB, Etzioni RD, Guerra CE, Perkins RB, Kondo KK. Screening for lung cancer: 2023 guideline update from the American Cancer Society. *CA: A Cancer Journal for Clinicians*. 2024 Jan;74(1):50-81.
22. Kumar A, Satheesha TY, Salvador BB, Mithileysh S, Ahmed ST. Augmented Intelligence enabled Deep Neural Networking (AuDNN) framework for skin cancer classification and prediction using multi-dimensional datasets on industrial IoT standards. *Microprocessors and Microsystems*. 2023 Mar 1; 97:104755.
23. Mohanty K, Subiksha S, Kirthika S, Sujal BH, Sokkanarayanan S, Bose P, Sathiyarayanan M. Opportunities of adopting AI-powered robotics to tackle COVID-19. In *2021 International Conference on COMMunication Systems & NETWORKS (COMSNETS)* 2021 Jan 5:703-708. IEEE. DOI:10.1109/COMSNETS51098.2021.9352917
24. Kavitha D, Murugan A, Sathiyarayanan M. Deep Analysis of Dementia Disorder Using Artificial Intelligence to Improve Healthcare Services. In *2021 International Conference on COMMunication Systems & NETWORKS (COMSNETS)* 2021 Jan 5:378-380. IEEE. DOI:10.1109/COMSNETS51098.2021.9352897.
25. Xu X, Li J, Zhu Z, Zhao L, Wang H, Song C, Chen Y, Zhao Q, Yang J, Pei Y. A Comprehensive Review on Synergy of Multi-Modal Data and AI Technologies in Medical Diagnosis. *Bioengineering*. 2024 Feb 25; 11(3):219.